

ION CHROMATOGRAPHY STANDARD REPORTING LIMITS



Curtis & Tompkins, Ltd.

Anions by EPA 300.0

CAS #	Compound	Water RL mg/L	Soil RL mg/Kg
24959-67-9	Bromide	0.2	2
16887-00-6	Chloride	0.2	2
16984-48-8	Fluoride	0.1	1
14797-55-8	Nitrate-N	0.05	0.5
14797-65-0	Nitrite-N	0.05	0.5
14808-79-8	Sulfate	0.5	5

Hexavalent Chromium by EPA 7199

CAS #	Compound	Water RL µg/L
18540-29-9	Hexavalent Chromium	0.5

Perchlorate by EPA 314.0

CAS #	Compound	Water RL µg/L
14797-73-0	Perchlorate	4